

TRIAFIX ARM 01

Arm support Ref: SRT 208

INSTRUCTIONS FOR USE

CE

Dear Customer!

Thank you for the confidence you have shown in purchaising Triamed products! We would like to inform you that Triamed products are made of the best performing materials, all having quality certificates and biocompatibility tests.

These orthoses comply with the Technical Specification and the requirements of Regulation (EU) MDR 2017/745 of the European Parliament on medical devices and comply with the requirements set out in Annex I - General Requirements for safety and performance- and Annex VIII - Classification rules. Conformity is declared on the basis of Article 52 (7), Annex II and Annex III of MDR 2017/745.

The orthosis is used according to the recommendations of the specialized physician.

Choosing the size - For optimal results, the appropriate size will be selected based on the measurements table of the product. A size too small may lead to intolerance to wearing or the damage of the product! A size too large significantly diminishes the efficacy of the orthosis.

Available sizes

Sizes	1	2	3
Forearm length (cm)	22-28	28-32	32-36
Reference	SRT208M1	SRT208M2	SRT208M3

The length of the forearm is measured

Indications

It is used to support the arm and forearm.

Description

The support is made of one single textile piece that supports the arm, provided with 2 adjustable support straps.

Hook and loop fastener.

The orthosis is an efficient support for the arm and forearm. Due to the two fastening straps, it does not allow the arm abduction and external rotation.

Composition

Three-layer laminated textile. The laminated textile ensemble behaves like a moisture regulator, ensuring a quick transfer of the moisture to outward

Positioning

The arm is positioned in the support at a 90° angle and the bands are fastened at the desired height.



To be washed by hand with warm water, no bleaching, no spinning, no ironing, no chemical cleaning!

Storage

At room temperature, in a dry space with good ventilation.

The legal guarantee of conformity according to Law no. 449/2003: 2 years.

Commercial warranty of the product: 30 days.

Date of production: see packaging label.

Validity date: no expiry date.

To validate the warranty, the seller must indicate the date of purchase. The buyer is obliged to keep and present the warranty coupon.

In order to solve the problems that appeared in the warranty, the following data must be notified: Article reference, batch and date of manufacture, indicated on the packaging box of the product.

Deadline for communication of the solution: 15 days.

Please send notifications and complaints to the following email address: suportclienti@triamed.ro

WARRANTY COUPON 30 days from the date of purchase	Batch and date of manufacture	
Product reference	Date of purchase	
The reason of complaint	Seller signature and stamp	

Manufacturer: TRIAMED SRL

România/Timiș/Timișoara, Str. Calea Buziașului nr.77, Hala 1-C1/A

Tel/Fax: 00 40 256 487388

office@triamed.ro, triamed@gmail.ro